

## Sputter coater SCD050 (Leica)

With the SCD050 samples for SEM can be sputter coated with Palladium/Gold to make them conductive.

Thickness can be monitored with a quartz thickness monitor.

Also carbon coating of samples or formvar coated grids (to make these hydrophilic) is possible.

